

# Temperature Compensation Chart

Because the temperature of the tested water will effect the result, you should adjust the *dip time* according the following chart:

°C / °F	Dip Time
0 / 32	40
1 / 34	38
2 / 36	36
3 / 37	34
4 / 39	32
5 / 41	30
6 / 43	28
7 / 45	27
8 / 46	26
9 / 48	25
10 / 50	24
11 / 52	23
12 / 54	22
13 / 55	20
14 / 57	19
15 / 59	18
16 / 60	17
17 / 62	16
18 / 64	15
19 / 66	14
20 / 68	13
21 / 70	12

°C / °F	Dip Time
22 / 72	11
23 / 73	10
24 / 75	10
25 / 77	10
26 / 79	9
27 / 81	9
28 / 82	9
29 / 84	8
30 / 86	8
31 / 88	8
32 / 90	8
33 / 91	8
34 / 93	7
35 / 95	7
36 / 97	7
37 / 99	7
38 / 100	7
39 / 102	7
40 / 104	6

**Note:** Measure the temperature of the water sample within  $\pm 2^{\circ}\text{C}$  /  $\pm 4^{\circ}\text{F}$